2025/11/29 17:10 1/3 PHD2 Guiding

PHD2 Guiding

C. Johnson

3/16/2018

updated Oct 3, 2018 by P Gabor

All of the required software is currently installed on the vattcontrol server; login as vattobs.

The legacy *VATT Guider App*only controls the stepper motors in the guide box.* Guiding is done with *PHD2*!

- (*) When using VATT4k and GUFI, these motors move the two filter wheels for the science camera, the guider camera's filter wheel, the guider camera's 3-axis translation stage, and the stage carrying the center mirror and the U mirror.
 - At startup
 - 1. start indiserver and correct drivers
 - click the indiserver icon on the desktop
 - (a terminal window should appear)
 - 2. start phd2 guiding
 - click the PHD2 icon on the desktop
 - (the PHD2 application should appear)
 - 3. configure phd2
 - 1. in the Main tool bar, click icon that looks like usb male connector
 - (connect equipment window should appear)
 - 2. in connect equipment window:
 - select camera type "INDI Camera",
 - select mount type "INDI Mount",
 - select aux type "None"
 - 3. in connect equipment window, in the camera row, click on the icon that looks like a screwdriver and wrench in an "X"
 - (INDI Configuration window should appear)
 - 4. in INDI Configuration window check following settings:
 - Hostname: localhost
 - port: 7600
 - driver: Apogee CCDDual CCD: Main

- 5. click the INDI button near the bottom of the window
 - (INDI options window should appear)
- 6. select the Apogee CCD tab
- 7. check following settings:
 - in the port section click the network button
 - in the network section:
 - subnet: 10.0.255.255
 - ip:port: 10.0.3.14:2571
- 8. click the connect button in the Connection field
 - (after a few seconds, a bunch of new fields and tabs should appear.)
- 9. near the bottom of the window, turn on the cooler
- 10. IF YOU ARE GOING TO BE AUTOGUIDING:
 - click on the "INDI-VATT-GUIDE" TAB
 - click "Connect"
 - (a UT clock should appear and begin incrementing.)
- 11. exit the INDI options window
- 12. click OK in the INDI Configuration window
- 13. click Connect in the Connect Equipment window, in the camera section
- 14. the button should now say Disconnect
- 15. click Close in the Connect Equipment window
- 16. In the Main tool bar (it may be positioned anywhere in the window),
 - click the Advanced setup (brain icon) button (second from the right); a new window should appear.
- 17. Set binning to the allowed maximum of 8×8:
 - Select the Camera tab,
 - in the group Camera-specific properties, set binning to 8 (pull down selector).
 - [Note. Ostensibly, binning can be set via INDI options as NxM where N and M may take any value you wish. This feature does not work. Binning can really be set only via the "brain" button.]
- 18. IF YOU WILL NOT BE AUTOGUIDING: Disable mount guide output:
 - Select the Guiding tab,
 - in the group Shared parameters disable mount guide output (uncheck the box); Guide output DISABLED should appear in the lower left of the main window. This should prevent PHD2 accidentally interfering, e.g., with the PEPSI guider.
- 19. Set up automatic frame capture:
 - Select the Global tab,
 - check Enable diagnostic image logging, then
 - in the group Save Guider Images check Until this count is reached, and
 - set the value (100 is the maximum).
 - The files are stored on the vattdev server!
 - The Dark Library is located in the directory /home/vattobs/.phd2/.
 - The directory /home/vattobs/PHD2/PHD2_CameraFrames... is automatically created for the whole session.
 - If you need to capture more than 100 images, make a renamed copy of the session directory; new fits files should start populating the original session directory.
- 4. Start Imaging

1. in the bottom left of the PHD2 window there is a button with 2 arrows forming a circle. Click that button.

2. the button should gray out, the stop sign button should turn red, and images should start appearing

From:

https://lavinia.as.arizona.edu/~tscopewiki/ - MOON

Permanent link:

Last update: 2018/10/03 15:39

